# **NEWS & VIEWS**

**Open Access** 

# Toward in vivo translation of superresolution localization photoacoustic computed tomography using liquid-state dyed droplets

Wonseok Choi<sup>1</sup> and Chulhong Kim<sup>1</sup>

Super-resolution localization photoacoustic computed tomography has been successfully performed in live animals using an injection of unprecedentedly absorptive liquid-state dyed droplets, which promises a disruptive modality for in vivo neuroimaging.

Photoacoustic (PA; or optoacoustic) tomography is a popular imaging modality that forms an image from the received ultrasound signals generated from targets that absorb light. PA tomography achieves both optical absorption contrast and high ultrasonic spatial resolution even in deep tissues. Due to this property, PA tomography can be implemented to see objects on scales ranging from nanometers to millimeters at depths up to several centimeters<sup>1</sup>. Among the diverse implementations of PA tomography, PA computed tomography (PACT) is a dedicated modality for imaging deep tissues (>4 mm) in live animals, healthy humans, and patients. PACT typically uses array transducers in various geometries to receive PA signals and reconstruct them into images via acoustic inversion or beamforming techniques. Since PA wavefronts propagate in spherical shapes, circular or spherical arrays are regarded as the best signal reception geometries<sup>2</sup>. In addition, advanced electronics and processing units can perform fast image reconstruction that enables near real-time imaging. With these features, PACT systems are capable of imaging the whole body of small animals or providing diagnostic images for clinical applications, such as images of the breast.

In PA tomography, there is generally a tradeoff relationship between spatial resolution and imaging depth, where the resolution is typically on the order of 1/200 of the desired imaging depth for biological tissues<sup>1</sup>. For instance, PACT can usually be applied at depths up to several centimeters with a relatively poor resolution of a few hundred micrometers, which is determined by the size of the acoustic parameters. In other words, while PACT is capable of imaging at deeper depths, it comes at the cost of not being able to visualize detailed structures of the tissue, which are important for fully understanding their basic functional mechanisms.

New super-resolution imaging techniques are breaking through the physical resolution limitations mentioned above. To date, two mainstream super-resolution imaging techniques have been proposed for PACT: localization imaging and fluctuation imaging. Localization imaging is analogous to photoactivation localization microscopy or stochastic optical reconstruction microscopy, and fluctuation imaging is analogous to super-resolution optical fluctuation imaging. Localization super-resolution imaging has been demonstrated in vitro with 2D and 3D PA images by Vilov et al. and Deán-Ben et al., respectively, sparsely distributed flowing using microbeads  $(10-30 \,\mu\text{m})^{3,4}$ . They were successful in distinguishing tubes with diameters of approximately four times smaller than the diffraction limit. Chaigne et al. used the superresolution fluctuation imaging technique, with PA signal fluctuation induced from varying speckle illumination and flowing absorbers, to calculate *n*-th order statistical values that enhanced the resolution by on the order of  $\sqrt{n^{5,6}}$ . While both techniques have been demonstrated successfully in vitro, there have been no publications describing

Correspondence: Chulhong Kim (chulhong@postech.edu)

<sup>&</sup>lt;sup>1</sup>Departments of Creative IT Engineering, Electrical Engineering, Mechanical Engineering, Interdisciplinary Bioscience and Bioengineering, Pohang University of Science and Technology, Pohang, Gyeongbuk 37673, Republic of Korea

<sup>©</sup> The Author(s) 2019

Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit http://creativecommons.org/licenses/by/4.0/.



in vivo applications yet. Potential reasons for the lack of in vivo applications include low signal contrast in deep tissues, motion artifacts due to a limited frame rate, and the lack of appropriate isolated particles that can flow stably through the narrow vascular network.

To overcome these limitations for in vivo superresolution imaging, Zhang et al. recently proposed dyed droplets as contrast agents for in vivo localization imaging<sup>7</sup>. Their key strategy is to use "liquid particles" that can smoothly flow through the blood vessels and generate prominent signal contrast, which is 300 times higher than the blood background. The dyed droplets were prepared by dissolving the IR-780 dye in oil and mixing the solution with water, which resulted in droplets with sizes ranging from 4 to  $30\,\mu\text{m}$ . For the in vivo demonstration in a mouse, the authors injected the dyed droplets through the left carotid artery of the mouse and used their ring-arraybased PACT system (Fig. 1a) to image the live mouse brain. The dyed droplets generated high PA signal contrast sufficient for the signal to be localized, thus enabling the reconstruction of a super-resolution image (Fig. 1b). As a result, they were able to produce images of the neurovasculature up to a depth of 4 mm with a sixfold improvement in the spatial resolution compared to the original PACT images (Fig. 1c, d)<sup>7</sup>.

While Zhang et al.'s proposal mitigated the low signal contrast challenge and met the requirement of smoothly flowing particles, the biggest hurdle of a low frame rate is yet to be overcome. The frame rates for PACT are limited by low laser repetition rates, which are generally approximately 5–20 Hz, needed for a high-power laser to achieve better signal-to-noise ratios in the images. Zhang et al. used a PACT system with a frame rate of 20 Hz, and they noted that the speed was not sufficient to capture even slowly flowing

droplets  $(1.3-7.5 \text{ mm/s})^7$ . Thus, there exists a need to achieve a high frame rate for a sufficient signal-to-noise ratio in deep tissue imaging using a laser with sufficient power. As mentioned in Zhang et al.'s paper, there is still room to improve the in vivo localization technique by using a dye with a higher absorption coefficient and regulating the droplet size. Therefore, once the low frame rate issue is addressed, it is expected that we will see more in vivo demonstrations of super-resolution PACT in the near future.

#### Acknowledgements

This work was supported by the National Research Foundation of Korea (NRF) grant funded by the Korea government (MSIT) (No. NRF-2019R1A2C2006269).

### Conflict of interest

C.K. has a financial interest in OPTICHO, which did not support this work. W.C. declares that he has no conflict of interest.

Received: 5 May 2019 Revised: 3 June 2019 Accepted: 13 June 2019 Published online: 25 June 2019

## References

- Wang, L. V. & Hu, S. Photoacoustic tomography: in vivo imaging from organelles to organs. *Science* 335, 1458–1462 (2012).
- Choi, W., Park, E., Jeon, S. & Kim, C. Clinical photoacoustic imaging platforms. Biomed. Eng. Lett. 8, 139–155 (2018).
- Vilov, S., Arnal, B. & Bossy, E. Overcoming the acoustic diffraction limit in photoacoustic imaging by the localization of flowing absorbers. *Opt. Lett.* 42, 4379–4382 (2017).
- Deán-Ben, X. L. & Razansky, D. Localization optoacoustic tomography. Light. Sci. Appl. 7, 18004 (2018).
- Chaigne, T. et al. Super-resolution photoacoustic fluctuation imaging with multiple speckle illumination. *Optica* 3, 54–57 (2016).
- Chaigne, T., Arnal, B., Vilov, S., Bossy, E. & Katz, O. Super-resolution photoacoustic imaging via flow-induced absorption fluctuations. *Optica* 4, 1397–1404 (2017).
- Zhang, P., Li, L., Lin, L., Shi, J. & Wang, L. V. In vivo super-resolution photoacoustic computed tomography by localization of single dyed droplets. *Light. Sci. Appl.* 8, 36 (2019).